



Corpus Christi – Nueces County Public Health
 District Laboratory
 1702 Horne Road
 Corpus Christi, TX 78416
 Office: (361) 826-7213
 Fax: (361) 826-7217

BAY WATER SAMPLES-BAHIA BAY

SPECIFIED LOCATIONS - SWIMMING/WATER CONTACT SPORT AREAS

DATE OF COLLECTION 05/19/11 BY Gary McAbee 361-727-9260
 TIME OF COLLECTION 10:00 a.m.
 DATE/TIME RECEIVED IN LABORATORY 05/19/11 @ 11:55 a.m. TEMP 2.0°C BY Diana Dominguez
 DATE/TIME REPORTED 05/20/11 @ 02:05 p.m. ANALYST Ken Diercouff/Estella Thomas

Sample Site	Chlorine Residual	Laboratory Number	*Fecal Coliform/100 ml	Comments
7. Windjammer Head	0.04	003103	15	
8. 206 Lands End	0.03	003104	2	
9. Windjammer ICW	0.04	003105	2	
10. 409 Square Rigger	0.09	003106	<1	
11. 504 Estes Drive	0.03	003107	2	
12. ICW Near Entrance	0.03	003108	<1	

Remarks: Test performed: IDEXX Membrane Filter (SM9222D) according to Standard Methods
 Form provided by the Corpus Christi-Nueces County Health Department Laboratory

*Texas Administrative Code: Surface Water Quality Standards for Contact Recreation:
 (a) Water samples from locations that exceed a measurement of 200 fecal coliforms/100ml must be recollected and submitted for analysis.